



EC Declaration of Conformity

(Directive 93/42/EEC), (Directive 2011/65/EU)

We,

Manufacturer's Name: Mortara Instrument, Inc.

Manufacturer's Address: 7865 North 86th Street
Milwaukee, WI 53224
USA

declare under our sole responsibility, that to the best of our knowledge, the product(s):

Product Name: Ambulo 2400 Ambulatory Blood Pressure Monitor

Part Number: ABP2400-XXX-XXXXX and ABP2400-XX
(X designates alpha characters denoting system configuration management codes important for post distribution servicing)

Product Options (Configurations): Hypertensive Diagnostic Software, Actigraphy, re-useable blood pressure cuffs (small, regular and large adult), EasyWear blood pressure cuffs and pouch, clip and strap, extension hose.

Lot and/or Serial Number(s): SN – "1YYWWXXXXXX"
(Where YY = 2 digit year;
WW = week code; and
XXXXXXX = unique number starting at 0000001)

GMDN Code and Term: 45617 Automatic-inflation electronic sphygmomanometer, portable, arm/wrist

Class (according to the criteria of Annex IX, 93/42/EEC): IIa (Rule 10)

*are in conformity with the dispositions of the directive which are applicable to them.
This declaration is based on the following elements:*

Directive 93/42/EEC: The ISO 13485 Certificate N° 7473 for approval of the quality system.
The EC Certificate ANNEX II N° 7472 for approval of the quality system.
Technical file (ref. Ambulo 2400, ANNEX VII) to demonstrate the conformity of the product to the essential requirements (ANNEX I).

Notified Body: LNE/G-MED (N° 0459)
1 rue Gaston Boissier
75724 Paris Cedex 15
France

European Union Representative: Mortara Instrument Europe, Srl
(European Headquarters, Italy)
Via Cimarosa 103-105
40033 Casalecchio di Reno, BO

Identification of Individual Signing/Location:


Mark Elliott
VP, Global QA/RA

Mortara Instrument, Inc.
7865 North 86TH Street
Milwaukee, WI 53224
USA

Date: April 14, 2017