

## **Corporate Headquarters**

4341 State Street Road, P.O. Box 220 Skaneateles Falls, NY 13153-0220 USA Phone: 800.535.6663 Fax: 315.685.3361 www.welchallyn.com



November 21, 2010

To Whom It May Concern:

The measurement time specification for the Braun PRO 4000 Thermometer may be further detailed as follows:

Measurement time:

i. Initial start-up time: 3-4 secondsii. Measurement time: 2-3 seconds

Sincerely,

Alisa R. Salibra

Global Category Manager, Thermometry

Welch Allyn